

National Textiles & Garment Quality Supervision Inspection Centre (ANHUI) Test Report

ADDRESS: No.25 ANQING
HEFEI CHINA

Tel: (0551) 85123869

Fax: (0551) 85125221

Postcode: 230000

NUMBER: YJ202006505E

DATE: 2020-03-20

APPLICANT: Anhui Hualing Textile Co., Ltd

SAMPLE DESCRIPTION: DISPOSABLE 3 layers MASK

ADDRESS: Delong road, industrial park, xindu town, tongcheng city, anhui province, China



ITEM	STANDARD (D GRADE)	RESULT	RATING
1. FILTER EFFICIENCY (SALT MEDIUM) (GB/T32610-2016)(%)			
-	≥90	94.9	PASS
2. PARTICLE PROTECTIVE PERFORMANCE (GB/T32610-2016)(%)			
-	≥65	79	PASS
3. RESPIRATORY RESISTANCE (GB/T32610-2016)(Pa)			
-INSPIRATORY RESISTANCE	≤175	76	PASS
-EXPIRATORY RESISTANCE	≤145	69	PASS

Remarks: 1. The protection effect grade is not marked on the sample mark, and it is judged according to the lowest standard requirement D grade.

2. The sample is not pretreated.

俞杰

PREPARED AND CHECKED BY

YU JIE

----END----